
T4x4

Floor Scale package



TRWS indicator



Top Quality Performer from Totalcomp

- 5,000 & 10,000 lb capacity
- 4'x4' & 5'x5' platforms
- Legal for Trade, NTEP
- Battery or AC operation
- Stainless steel indicator
- Checkered safety tread deck

email: sales@totalcomp.com



Specifications

Heavy duty floor scale package includes a TRWS indicator and floor scale base.

The TRWS indicator has a LCD display and keys for UNIT, PRINT, G/N, TARE, ZERO, ON/OFF. Housed in a stainless steel case. Internal rechargeable battery for 35 hours of use. 120 VAC adapter for continuous use. Indicator is NTEP 13-091 III, 5,000 div Legal for Trade.

The rugged base is sturdy with channel construction, Checkered safety tread deck. Painted steel. Comes with 15 ft of shielded load cell cable to connect to the indicator.

Base is NTEP 16-016 III, 5,000 div Legal for Trade

Part Number	Capacity	Platform	Weight	Stock
T4x4-5K Package	5,000 lb x 1 lb	4' x 4'	260 lb	4737
T4x4-10K Package	10,000 lb x 2 lb	4' x 4'	260 lb	4738
T5x5-5K Package	5,000 lb x 1 lb	5' x 5'	345 lb	4768
T5x5-10K Package	10,000 lb x 2 lb	5' x 5'	345 lb	4769

Electrical Power source

120V AC adapter 12VDC, 500ma Internal lead acid rechargeable battery Temperature
Compensated Range
-10 deg C to +40 deg C
Operating Range

-16 deg C to +65 deg C

Dimensions

Indicator

7.75"Wx3.25"Dx4.63"H

email: sales@totalcomp.com

Base

4'x4'x3.75"H 5'x5'x3.75"H

Totalcomp Limited Warranty

Totalcomp warrants this product to be free of defects in materials and/or workmanship and suitable for the purpose intended as outlined in this sheet. This warranty is effective and shall cover the purchaser for one year from the date of shipment from our plant. If this product is found to be defective by our inspection in accordance with the above listed criteria, we will replace or repair it at our expense. For warranty service, please obtain a return authorization number from us and return the item, shipping prepaid, with a written description of the problem. We will respond promptly with the results of our evaluation.

